

Ahmed® Glaucoma Valve



	Ahmed® Glaucoma Valve Model FP7	Ahmed® Glaucoma Valve Model FP8
Type	Valve	Valve
Surface Area	184 mm ²	102 mm ²
Plate Length	16.0 mm	11.0 mm
Plate Width	13.0 mm	11.0 mm
Plate Thickness	2.1 mm	2.1 mm
Tube Length	25.4 mm	25.4 mm
Plate Fenestrations	3	0
Material	Medical grade silicone	Medical grade silicone

Reimbursement

CPT® Code	Description of Procedure
66179	Aqueous shunt to extraocular reservoir; without patch graft
66180	Aqueous shunt to extraocular reservoir; with patch graft

Ordering Information

Ahmed® Glaucoma Valve, Model FP7	FP7
Ahmed® Glaucoma Valve, Model FP8	FP8



Ready to get started?

Join LEARN, our online training platform, to instantly gain access to product education and training modules

www.learn.newworldmedical.com

For more information, contact your local sales representative or email us at communications@newworldmedical.com

Humanity's Vision Is Our Focus



References:

- ¹ Market Scope. 2018 Glaucoma Surgical Device Report: A Global Market Analysis for 2017 to 2023. St. Louis, MO: Market Scope, LLC; August 2018.
- ² American Academy of Ophthalmology. Preferred Practice Pattern®: Primary Open-Angle Glaucoma. San Francisco, CA: American Academy of Ophthalmology; January 2016. [https://www.aaojournal.org/article/S0161-6420\(15\)01276-2/fulltext](https://www.aaojournal.org/article/S0161-6420(15)01276-2/fulltext) Accessed October 5, 2018.
- ³ Budenz DL, Barton K, Gedde SJ, et al. Five-year treatment outcomes in the Ahmed Baerveldt comparison study. *Ophthalmology*. 2015;122(2):308-316.
- ⁴ Budenz DL, Feuer WJ, Barton K, Schiffman J, Costa VP, Godfrey DG, Buys YM; Ahmed Baerveldt Comparison Study Group. Postoperative Complications in the Ahmed Baerveldt Comparison Study During Five Years of Follow-up. *Am J Ophthalmol*. 2016 Mar;163:75-82
- ⁵ Law SK, Kormmann HL, Giaconi JA, Kwong A, Tran E, Caprioli J. Early aqueous suppressant therapy on hypertensive phase following glaucoma drainage device procedure: a randomized prospective trial. *J Glaucoma*. 2016;25(3):248-257.